



Biosafety Level 2	
SOP Number:	175.01
Date:	17-June-2020

Biosafety Level 2

1. General

- 1.1 The static magnetic field in the CFMM MRI facilities is always present. It is essential that everyone entering the facility is aware of the presence of the magnetic field, since we cannot otherwise detect it (i.e. magnetic fields cannot be seen or felt).
- 1.2 Biosafety Level 2 represents pathogens that can cause human or animal disease, but, under normal circumstances, are unlikely to be a serious hazard to healthy laboratory workers, the community, or the environment.
- 1.3 Any Level 2 studies requesting MRI time, must have an Animal Use Protocol (AUP) or REB approval in place. This document is to be used in conjunction with approved procedures outlined in the AUP or REB protocols.
- 1.4 Signage indicating “Authorized Personnel Only – Proper PPE Required” or a similar warning statements, is posted during scans of these nature. A physical barrier is also utilized in the 3T/7T area to prevent unauthorized personnel from entering during these studies.
- 1.5 Cleaning and decontamination after these experiments will be carried as per SOP 170 “Cleaning, Waste Disposal and Chemical Storage”.



Biosafety Level 2	
SOP Number:	175.01
Date:	17-June-2020

Revision Chronology		
Version Number	Date	Changes
175.01	17 June 2020	First Version (MB)

CFMM Director Signature:

Date:

17 – June - 2020
